

Item number 290407

Excellent

Long examination gloves, Accelerator-free, Abena Excellent, M, blue, nitrile, powder free, accelerator free

- ✓ Long cuff
- ✓ Elastic
- ✓ Fits to hand
- ✓ Accelerator free

MD CE CAT III

Product Description

Long examination glove, no rubber chemicals/accelerators. Can be used for a variety of tasks such as cleaning, care, food handling, laboratory, medication handling, cytostatics and hormone creams. Contains neither latex proteins nor accelerators. Safe for use in the event of latex allergy and contact allergy to rubber chemicals.

300 mm · MEDIUM GRIP · LATEX · 374-1:2016/Type C K · EN 420 EN 455 · AQL 1.5 · EN374-5:2016 VIRUS

Specifications

Base name	Long examination gloves, Accelerator-free
Brand	Abena
Sub-brand	Excellent
Size	M
Properties	Latex free, finger textured, medium grip, rolled cuff, long cuff, accelerator free
Features	Powder free, accelerator free
Single or multiple use	Single use
Material	Nitrile
Technical ingredients	Nitrile NBR, Titanium Dioxide, Zinc oxide, Carbon Dioxide, Potassium Hydroxide, Pigment
Length/depth	300 mm
Width	95 mm
Thickness	0,06 mm
Certifications	CE, Food contact materials, CAT III, MD
CE Category (Personal Protective Equipment)	CAT III
CE Class (Medical Devices)	Class I
Product or test standards	EN 455, EN 374, EN 420, AQL 1.5, EN 374-1:2016 Type C K, EN 374-5:2016 VIRUS
Directives, regulations and acts	MDR (EU) 2017/745, 425/2016/EU, 10/2011/EU
Because of the products composition it has an expiry date from the production date on:	3 years
Storage Instructions	Store dry, clean and at room temperature.
Product Disposal Instructions	Sort as plastic. If contaminated dispose off as clinical waste.

Instructions for use/application

Examine the gloves for flaws and defects before use.

Abena · Tel. +45 74 31 18 18 · info@abena.com · www.abena.com

Packaging data

Unit	Contains	Length	Width	Height	EAN
cll	1000 pcs	36.5 cm	26 cm	27 cm	5703538153822
pck	100 pcs	26 cm	12.6 cm	7 cm	5703538153808
pair	2 pcs	0 cm	0 cm	0 cm	

MD

Regulation (EU) 2017/745 is a regulation of the European Union on the clinical investigation and sale of medical devices for human use. It repeals Directive 93/42/EEC, which concerns medical devices, and Directive 90/385/EEC, which concerns active implantable medical devices, on 26 May 2021.

The glass-fork-symbol guarantees that products live up to the demands stated in the European legislation for food contact materials. The symbol is not mandatory but advised on products used for food contact.

CE

The CE mark guarantees that a product is safe to use and complies with all safety precautions. CE stands for Conformité Européenne (European Conformity) and is mainly found on electronic equipment, safety equipment, construction products, and medical devices.

CAT III

Third-party type approval is required for all Category III personal protective equipment (PPE) products. CAT III products will be accompanied by a notified body number for the 3rd party verification of the product's safety.

MEDIUM GRIP

The glove has a medium grip.



The glove is 300 mm long.



The product does not contain latex.



The glove has a finger textured surface.

EN 455

EN 455 consists of four standards that gloves must be tested against in order to be a medical glove. This standard specifies the requirements for disposable medical gloves.

EN 420

The standard specifies the requirements for protective gloves in relation to general requirements and test methods.



The standard specifies the requirements for protective gloves against hazardous chemicals and microorganisms. This part of the standard describes the glove's resistance to bacteria, fungi and viruses.



The standard specifies the requirements for protective gloves against chemicals and microorganisms. Type C has been tested for a permeation time of at least 10 minutes with the chemical sodium hydroxide.


Gloves | Nitrile
POWDER-FREE | MEDICAL EXAMINATION GLOVES
M
7-8
Excellent
LONG CUFF | ACCELERATOR FREE
100 pcs.